

# Japan Charged-Particle Nuclear Reaction Data Group

Department of Physics, Faculty of Science  
Hokkaido University  
060-0810 Sapporo, JAPAN

E-mail: services@jcprg.org  
Internet: http://www.jcprg.org/

Telephone +81(JPN)-11-706-2684  
Facsimile +81(JPN)-11-706-4850

## Memo CP-E/116 (Revised)

**Date:** June 6, 2007  
**To:** Distribution  
**From:** OTSUKA Naohiko  
**Subject:** Reference frame of analyzing power ~~and Rutherford ratio~~

Angular distribution and analyzing power are related to count numbers as follows:

$$\sigma(\theta) = C \Delta N(\theta) / \Delta \Omega(\theta)$$
$$pA_y(\theta) = (\Delta N_U(\theta) - \Delta N_D(\theta)) / (\Delta N_U(\theta) + \Delta N_D(\theta)),$$

where  $C$  is a constant,  $\Delta N(\theta)$ ,  $\Delta N_U(\theta)$ ,  $\Delta N_D(\theta)$  are count numbers for unpolarized beam, beam polarized along  $+y$  and beam polarized along  $-y$  axis at a detector having solid angle  $\Delta \Omega(\theta)$ .  $p$  is beam polarization. The setting angle of the detector is  $\theta_{\text{lab}}$  in the laboratory system and it is  $\theta_{\text{cm}}$  in the center of mass system. Two angles in reaction  $A(a,b)B$  are related by the relation:

$$\tan \theta_{\text{lab}} = \frac{\sin \theta_{\text{cm}}}{\gamma + \cos \theta_{\text{cm}}}, \text{ where } \gamma = \sqrt{\frac{m_a m_b}{m_A m_B} \frac{E_{\text{cm}}}{E_{\text{cm}} + Q}} \text{ and } E_{\text{cm}} = \frac{m_A}{m_a + m_A} E_{\text{lab}}$$

All count numbers does *not* depend on the reference frame (=frame where measurement is done).  $\Delta N_{\text{lab}}(\theta) = \Delta N_{\text{cm}}(\theta)$  etc., but the solid angle depends on the reference frame:  $\Delta \Omega_{\text{lab}}(\theta) = J \Delta \Omega_{\text{cm}}(\theta)$ , where  $J$  for reaction is

$$J = \frac{|1 + \gamma \cos \theta_{\text{cm}}|}{(1 + \gamma^2 + 2\gamma \cos \theta_{\text{cm}})^{3/2}}.$$

Therefore angular distribution at the detector depends on the reference frame, i.e.  $\sigma_{\text{lab}}(\theta) = \sigma_{\text{cm}}(\theta) / J$ , while analyzing power at the detector does *not* depend on the reference frame, i.e.  $A_{y \text{ lab}}(\theta) = A_{y \text{ cm}}(\theta)$ . This means DATA-CM is useless for analyzing power. It is same for angular distribution of other polarization quantities. ~~We can prove that Rutherford ratio also takes same value in the laboratory and center of mass system by same way.~~

Below I summarize the usage of data heading for angular distribution, ~~Rutherford ratio~~ and analyzing power:

Angle (x)	Quantity (y)	Heading (x)	Heading (y)
$\theta_{lab}$	lab	ANG	DATA
$\theta_{lab}$	cm	ANG	DATA-CM
$\theta_{cm}$	lab	ANG-CM	DATA
$\theta_{cm}$	cm	ANG-CM	DATA-CM
<del><math>\theta_{lab}</math></del>	<del><math>\frac{1}{r}</math></del>	<del>ANG</del>	<del>DATA</del>
<del><math>\theta_{cm}</math></del>		<del>ANG-CM</del>	<del>DATA</del>
$\theta_{lab}$	$A_y$	ANG	DATA
$\theta_{cm}$		ANG-CM	<b>DATA</b>

The reference frame of quantity is often omitted by authors, and JCPRG usually assumes the reference frame of the quantity given and reference frame of the angle are same for such data.

### Distribution:

S. Babykina, CAJaD	A. Blokhin, CJD	J.H. Chang, KAERI	M. Chiba, JCPRG
F.E. Chukreev, CAJaD	S. Dunaeva, NDS	S. Ganesan, BARC	Z.G. Ge, CNDC
O. Gritzay, KINR	A. Hasegawa, NEA-DB	H. Henriksson, NEA-DB	A. Kaltchenko, KINR
J. Katakura, JAEA	K. Katō, JCPRG	Y.O. Lee, KAERI	S. Maev, CJD
V.N. Manokhin, CJD	V. McLane, NNDC	A. Mengoni NDS	M. Mikhaylyukova, CJD
A. Nichols, NDS	C. Nordborg, NEA-DB	P. Obložinský, NNDC	Y. Ohbayasi, JCPRG
A. Ohnishi, JCPRG	N. Otuka, JCPRG	V. Pronyaev, CJD	D. Rochman, NNDC
O. Schwerer, NDS	S. Tákacs, ATOMKI	S. Taova, VNIIEF	T. Tárkányi, ATOMKI
V. Varlamov, CDFE	M. Vlasov, KINR	M. Wirtz, NDS	H.W. Yu, CNDC
V. Zerkin, NDS	Y.X. Zhuang, CNDC	EXFOR, NEA-DB	